

RIFE RAY MACHINE NO. 4

NAME	OSCILLATOR		FREQUENCY KC	GROUP 1		GROUP 2	
	S	DIAL		S	DIAL	S	DIAL
BX Filterpassing	3	85.50	1604 = 1,604,000 Hertz	18.8	6	21.0	
TYPHOID Filterpassing	3	76.66	1445 = 1,445,000 Hertz	23.2	6	26.6	
TYPHOID Rod	3	35.00	760 = 760,000 Hertz	76.2	6	79.0	
ACTINOMYCOSIS (Streptothrix)	4	20.75	192 = 192,000 Hertz		8	77.1	
STAPHYLOCOCCUS	4	85.25	478 = 478,000 Hertz	27.2	7	44.5	
B. COLI Rod	4	73.50	417 = 417,000 Hertz	42.6	7	62.5	
DIPLOCOCCUS PNEUMONIAE	4	75.33	427 = 427,000 Hertz	40.0	7	59.1	
BACILLUS TETANI (Tetanus)	4	36.5	234 = 234,000 Hertz		8	49.25	
STREPTOCOCCUS PYOGENOUS	3	31.00	720 = 720,000 Hertz	82.2	6	86.2	
BACILLUS TUBERCULOSIS Rod	4	64.50	369 = 369,000 Hertz	57.7	7	80.6	
B Coli-fil	3	36	770 = 770,000 Hertz	74.5	6	77.25	
B Anthrax	5	81.26	139.2 = 139,200 Hertz		9	29.1	
Treponema Paludium	3	37.25	789 = 789,000 Hertz	71.75	6	74	
Staphylococcus	4	36	233 = 233,000 Hertz		8	49.5	

RIFE RAY MACHINE NO. 4

<u>GROUP 1</u>			<u>GROUP 2</u>		
<u>Switch</u>	<u>From KC</u>	<u>To KC</u>	<u>Switch</u>	<u>From KC</u>	<u>To KC</u>
1			1		
2	22,500.	15,400.	2		
3	16,300.	9,270.	3	15,600.	9,350.
4	9,500.	4,800.	4	9,370.	4,640.
5	4,850.	2,150.	5	4,730.	2,130.
6	2,175.	615.	6	2,140.	650.
7	615.	282.	7	670.	325.
8			8	332.	168.
9			9	172.	87.
10			10		

No setting between 650 KC AND 2130^{KC} that works smoothly on W.M.A.
 800 KC. about 156 D 70 3.256 D 72.5

Test of No. 4 Rife Ray Machine
Rife Laboratory, San Diego

On October 4th, 1935, Dr. Milbank Johnson arrived at the laboratory with Philip Hoyland of Pasadena to make the final tests on the new machine Mr. Hoyland had just completed and it was designated as Rife Ray Machine No. 4.

The equipment used in the test is as follows:

Seventy-three (73) pound chunk of the rump of a freshly killed horse containing the femur intact.

Specially constructed glass ampules having eye blown in one side and pipet drawn out opposite, each being numbered.

Lengths of silk fish line with tags attached.

Eighteen inch lance with hole in blade, for threading line.

Test tube of known B. Typhosis in broth culture media.

Pipet for filling ampules.

Petrographic microscope at 1400X.

Slides, cover glasses, etc.

Petri dishes containing agar media.

No. 4 Rife Ray Machine with quartz tube.

Dr. Johnson had devised a very unique method for testing the MOR of different known cultures. The ampules, eight in number, were half filled with B. Typhosis culture in broth media and sealed, numbered and tagged with silk fish line of suitable length. The chunk of horse meat was placed on a bench 10 cm under the quartz tube attached to the No.4 Rife Ray Machine. The lance was threaded with the line

attached to ampule No. 2 and run through the chunk of horse meat about 3 cm from the top; the lance unthreaded while in this position was withdrawn and the attached ampule drawn into the incision by means of the silk fish line.

Ampule No. 3 was placed at a depth of 6 cm in the same manner.

" No. 4 " " " " " 9 cm " " " "

" No. 5 " " " " " 12 cm and under the femur.

" No. 6 " " " " " 15 cm

" No. 7 " " under the whole chunk of horse meat.

" No. 9 (No. 8 having been broken) was placed on top of the whole chunk of horse meat.

Ampule No. 1 was isolated in a distant room as a control.

The test being ready, the No. 4 Rife Ray Machine was set for the MOR of B. Typhosis (rod) and also for the filter-passing form. It was turned on for an exposure of 3 minutes/after which the ampules were withdrawn, opened, and slides prepared, from each one.

Slides upon being examined were found to contain no motile rods, and to be thoroughly agglutinated and seemingly shattered. On the slide prepared from Control No. 1 Ampule, rods were found very active. Culture prepared from all ampules in petri dishes and replaced and left in incubators 48 hours showed no growth and no motility. Neither growth nor motility showed in ampules which had all been resealed and placed in incubator for 48 hours. Culture from Control Ampule No. 1 (not exposed) showed good growth and high motility in both cultures. Also grew well on agar plate. It appears that the No. 4 Rife Ray Machine is OK.

This test witnessed and written
by Jack Free, asst. to Dr. R.R.Rife